



# MEIJI TECHNO CO.,LTD.

322-1, Chikumazawa, Miyoshi-machi, Iruma-gun  
Saitama 354-0043, Japan

Tel : 81-(0)49-259-0111  
Fax : 81-(0)49-259-0113  
E-mail : meiji@meijitechno.co.jp  
<http://www.meijitechno.com>



## RoHS Declaration of Conformity

**06/08/2011**

**MEIJI TECHNO CO., LTD.**

**322-1 Chikumazawa, Miyoshi-machi, Iruma-gun Saitama Japan**

To Whom It May Concern

We, Meiji Techno America do hereby declare under our sole responsibility, declare under sole responsibility that all models of MEIJI Brand microscope to which this declaration is in conformity with Directive 2011/65/EU of the European Parliament and of the Council of June 8<sup>th</sup> 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

European standard requirements: 2011/65/EU Restriction of Hazardous Substances

### Content Summary

Lead (Pb)	< 0.1% or 1000 ppm
Mercury (Hg)	< 0.1% or 1000 ppm
Cadmium (Cd)	< 0.01% or 100 ppm
Hexavalent chromium (Cr+6)	< 0.1% or 1000 ppm
Polybrominated biphenyls (PBB)	< 0.1% or 1000 ppm
Polybrominated diphenylethers (PBDE)	< 0.1% or 1000 ppm

### Products List

All Meiji Brand Microscope

Sincerely,

Mrs. Sato, President  
Meiji Techno America., Ltd.  
322-1, Chikumazawa  
Miyoshi-machi, Iruma-gun,  
Saitama 354-0043 Japan